

17 Harrison Street  
Marshalltown

P O Box 61542  
Marshalltown 2107  
Tel : (011) 688-1400 Fax : (011) 688-1556



Johannesburg Water

**REQUEST FOR QUOTATION**

INITIATING DEPARTMENT	INITIATOR
CYDNA -ONLINE	Desegan
QUOTATION REFERENCE	COLLECTIVE NO.
RFQJW359NM-CHLORINE ANALYSER	
QUOTATION REQUESTED FROM	

QUOTATION VALIDITY
60 days

PAGE NO.
CLOSING DATE AND TIME
<b>4-Nov-20</b> 16:00
Date of Issue
<b>28 October 2020</b>
VALIDITY
: DAYS

**QUOTATIONS WILL BE EVALUATED ON THE 80/20 POINT SCORING SYSTEM.80 POINTS WILL BE ALLOCATED TO PRICE AND THE REMAINING 20 POINTS WILL BE ALLOCATED FOR BBBEE AND PREFERENTIAL PROCUREMENT ALL SUPPLIERS RESPONDING TO QUOTATIONS SHOULD BE REGISTERED ON CENTRAL SUPPLIER DATABASE(CSD)**

**JW CONTACT PERSON:Nompumezo,e-mail address:nompumezo.mramba@jwater.co.za.....**

**TELEPHONE NUMBER:011 688 1656,FAX NO:086 611 8167**

ITEM NO.	DESCRIPTION	DESCRIPTION OF ITEM OFFERED	UOM	QTY REQUIRED	PRICE QUOTED EXCL. OF V.A.T.	DIS
	<b>SPECIFICATIONS FOR RESIDUAL CHLORINE ANALYSER.</b>					
	Range: 0-5mg/l					
	Accuracy: +-55 or +-0.03mg/l (ppm) as CL2					
	Resolution: 0.01mg/l (ppm)					
	Lower Limit of Detection: 0.03mg/l (ppm)					
	Cycle Time: Approx.2 minutes					
	Operating Temperature range: 5 - 40C					
	Output signal: 4 -20 Ma					
	Mounting style: Wall mount					
	Light Source: Class 1 LED					
	Weight: 11 lbs (5kg) - 23.13 lbs (10.49)					
	Display: Blacklit LCD					
	Enclosure: ABS plastic					
	Power Requirements: 100 - 240 V AC / 47 -63 Hz / 2.5 A					
	Sample flow rate: Approx. 150-500 ml per minute					
	<b>NOTE: Training of 10 personnel (divided into 2 groups) will need to be conducted upon commissioning of the equipment.</b>					
	<b>Quotations to: Nompumezo.mramba@jwater.co.za</b>					
	<b>For more information contact Shumani on 011 483 9500</b>					
	<b>NB:All suppliers responding to Rfq's should use their own company letter head not Jw Rfq Template AND MAKE SURE THEIR EMAIL ADDRESS IS VISIBLE ON THEIR QUOTATION.</b>					
	<b>QUOTATION REFERENCE AS ABOVE JW....</b>					

OFFICIAL STAMP		<b>SUPPLIER DETAILS</b> AUTHORISED BY:..... SIGNATURE:..... DATE:.....		<b>CONDITIONS</b> 1. QUOTATIONS RECEIVED AFTER CLOSE OF BUSINESS ON THE CLOSING DATE WILL NOT BE ACCEPTED. 2. QUOTATIONS WITHOUT BRAND NAMES WHERE REQUIRED WILL NOT BE ACCEPTED 3. PRICES QUOTED MUST BE AS PER THE UNIT INDICATED AND BE EXCLUDED OF VAT 4 QUOTATIONS WITHOUT THE SUPPLIER'S AUTHORISED SIGNATURE WILL NOT BE ACCEPTED. 5 ACCEPTANCE OF A QUOTATION WILL BE SUBJECT TO JOHANNESBURG WATER'S SUPPLY CHAIN POLICY		
----------------	--	---	--	--	--	--